

AB Tip İnceleme Sertifikası **EU Type-Examination Certificate**

Belge No / Certificate No

: 162-21-01-R01

Belgelendirme Tarihi - Bir Sonraki Belge Tarihi /

Certification Date / Certificate Validity Date

: 12.03.2021-16.02.2026

Belge Gecerlilik Tarihi / Document Validity Period: 5 yıl / 5 years

Firma Unvanı ve Adresi /

Company Name and Address

: ASTEK METAL MEDİKAL YAPI

TEKNOLOJÍ SAN. TÍC. LTD. ŞTÍ.

Akçaburgaz Mah. 3032. Sk. No: 4/1

Esenyurt / İSTANBUL

Ürün Adı /Modeller / Product Name / Models

Direktifi / Directive

Modülü/Kategori / Module / Category

: ARMED AR PRO 2

: 2016/425 REGULATION

: B MODÜLÜ/ KATEGORİ III

MODULE B / CATEGORY III

: MNA M-2020-00799, M-2021-00316

Test Rapor No/ları / Test Report No

Ürün Tipi / Product Type:

EN 149:2001+ A1:2009 Solunumla ilgili koruyucu cihazlar - Parçacıklara karşı koruma amaçlı filtreli yarım maskeler/ Respiratory protective devices - Filtering half masks to protect against particles

Ürünün Malzeme Bilgisi / Product Material Information: ARMED AR PRO 2 model ürünleri kumaş, elastik kayış, burun klipsi, filtre katmanı kullanılarak imal edilmiştir./ ARMED AR PRO 2 model products are manufactured using fabric, elastic strap, nose clip, filter layer.

Revizyon nedeni/ Reason for revision: Farklı renkte ürünler eklenmiştir/ Different color products have been added.

Volkan AKIN 12.03.2021 Karar Verici / Approver

Okan AKEL 12.03.2021

Sirket Müdürü / General



MNA Laboratuvarları San. Tic.Ltd .Şti Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/İstanbul Tel: 0216 574 07 08 Faks: 0216 575 13 31 www.mnalab.com

U-Form-002/Rev.04/12.03.2020




ATTACHMENTS (162-21-01-R01)

To certify the PPE product at Category III level, C2 or D module is accompanied by applying one of the conformity assessment methods along with the EU Type Examination (Module B).

Model: ARMED AR PRO 2

PPE SPECIFICATION	PERFORMANCE LEVELS
Classification	FFP2
Reusable / Single Shift Use	NR


PPE produced as a single unit to fit an individual user, all the necessary instructions for manufacturing such PPE on the basis of the approved basic model:



MNA LABORATORIES SAN. TIC. LTD. ŞTİ declares that the above-mentioned product meets the requirements of the directive according to the EU Directive 2016/425, the safety of the product is covered by the conditions and use specified in this certificate and in the technical file.



ATTACHMENTS (162-21-01-R01)



MNA Laboratuvarları San. Tic.Ltd .Şti Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/ İstanbul Tel: 0216 574 07 08 Faks: 0216 575 13 31 www.mnalab.com



ATTACHMENTS (162-21-01-R01)



DOCUMENTS IN THE TECHNICAL FILE

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- Technical Report



TECHNICAL EVALUATION REPORT (162-21-01-R01)

Report No : 162-21-01-R01

Report Date :12.03.2021

Application No : 162-21-01

1. COMPANY INFORMATION:

ASTEK METAL MEDİKAL YAPI TEKNOLOJİ SAN. TİC. LTD. ŞTİ. Akçaburgaz Mah. 3032. Sk. No: 4/1 Esenyurt / İSTANBUL

Tel: +90 212 219 72 58

2. PPE INFORMATION:

Disposable and non-sterile half mask made of particulate protection fitler material.


3. PPE TYPE IDENTIFICATION

EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles - Requirements, testing, marking

4. PPE PICTURES












TECHNICAL EVALUATION REPORT (162-21-01-R01)



ARMED AR PRO 2

5. PPE DIMENSIONS:

ARMED AR PRO 2 model has been found to be produced using standart sizes.

6. PPE PRODUCT MATERIAL INFORMATION:

The product is made of elastic strap, nonwoven fabric on the outer and inner layers and fitler material on the middle layer.

7. ESSENTIAL HEALTH AND SAFETY REQUIREMENTS

- A visual inspection was made according to EN 149:2001 +A1:2009 for ergonomics.
- · Protection levels and degrees are defined by the manufacturer.
- Suitable construction materials were determined by visual inspection according to EN 149:2001 +A1:2009.

8. ANALYSIS AND EVALUATIONS:

EN 149:2001 +A1:2009

TESTS	PARAMETER PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION	
		FFP1	FFP2	FFP3			
Part 7.3 Visual inspection	Shall also the marki supplied by the man	(T)	facturer		Appropriate	-	PASS
Banned Azo Dyes	< 30 mg/kg) mg/kg			< 5 mg/kg	< 30 mg/kg	PASS
Part 7.4 Packaging	Particle filtering half mask shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.			Appropriate	-	PASS	
Part 7.5 Material	When conditioned in accordance 8.3.1 & 8.3.2 the particle filter half mask shall not collapse.				Appropriate	-	PASS
Part 7.6 Cleaning and disinfecting	After cleaning and di particle filtering half penetration require class.	f mask sh	nall sati	sfy the	Not applicable	-	Not applicable



TECHNICAL EVALUATION REPORT (162-21-01-R01)

Part 7.7	No negative comments should be made by	Appropriate	-	PASS
Practical performance	the test subject regarding any of the criteria evaluated.			
Part 7.8 Finish of parts	Parts of the device likely to come into contact with the wearer shall have no sharp edge or burrs.	Appropriate	-	PASS

TESTS	PARAMETER PERFORMANCE LEVELS		RESULTS	PERFORMANCE LEVELS	EVALUATION		
		FFP1	FFP2	FFP3			
Part 7.9.1 Total inward leakage	At least 46 out of the 50 individual exercise result	<25	<11	<5	See the table below	FFP2	PASS
	At least 8 out of the 10 individual wearer arithmetic means	<22	<8	<2	See the table below	FFP2	PASS

Total Inward Leakage (%)											
	Exercise 1	Exercise 2	Exercise 3	Exercise 4	Exercise 5	Average					
Subject 1 (As recieved)	8.1	7.3	7.9	6.6	7.9	7.6					
Subject 2 (As recieved)	7.5	7.2	7.8	7.5	7.3	7.5					
Subject 3 (As recieved)	8.3	7.9	8.3	5.2	7.0	7.3					
Subject 4 (As recieved)	6.7	7.7	7.5	8.1	6.3	7.3					
Subject 5 (As recieved)	8.1	7.5	7.7	7.4	6.7	7.5					
Subject 6 (After temperature conditioning)	7.7	7.1	8.4	7.6	7.4	7.6					
Subject 7 (After temperature conditioning)	7.6	7.8	7.5	7.8	7.5	7.6					
Subject 8 (After temperature conditioning)	7.6	7.5	7.8	7.0	7.4	7.5					
Subject 9 (After temperature conditioning)	7.7	7.5	7.8	8.5	7.6	7.8					
Subject 10 (After temperature conditioning)	8.5	7.0	6.4	7.5	7.1	7.3					

Subject facial dimensions

Subject	Face Length (mm)	Face Width (mm)	Face Depth (mm)	Mouth Width (mm)	
1	133	132	132	65	
2	125	144	116	67	
3	126	135	124	75	
4	123	133	134	74 73 66	
5	117	135	122		
6	122	142	133		
7	113	132	114	75	
8	135	123	123	65	
9	122	135	133	74	
10	135	142	125	83	



TECHNICAL EVALUATION REPORT (162-21-01-R01)

TESTS	PARAMETER	PERFORMANCE LEVELS		RESULTS	PERFORMANCE LEVELS	EVALUATION	
		FFP1	FFP2	FFP3			
Part 7.9.2 Penetration of filter	Sodium chloride, 95 L/min %, max	% 20	% 6	%1	See the table below	FFP2	PASS
material	Paraffin oil, 95 L/min %, max	% 20	% 6	% 1	See the table below	FFP2	PASS

Penetration of filter material	Sodium Chloride (%)	Paraffin Oil (%)	
As recieved	3.5	3.3	
As recieved	2.2	2.4	
As recieved	2.7	3.2	
After the simulated wearing treatment	2.4	3.1	
After the simulated wearing treatment	2.6	2.6	
After the simulated wearing treatment	2.1	3.8	
Mechanical strength and temperature conditioning	4.6	4.8	
Mechanical strength and temperature conditioning	4.4	4.3	
Mechanical strength and temperature conditioning	4.3	4.7	

TESTS PARAMETER	PARAMETER	PERFO	RMANO	E LEVELS	RESULTS	PERFORMANCE	EVALUATION
		FFP1	FFP2	FFP3	1	LEVELS	
Part 7.10 Compatibility with skin		aterials shall not be known to be likely to use irritation or any other adverse effect to ealth			Appropriate	-	PASS
Part 7.11 Flammibility	Mask shall not burn o for more than 5 s	r not to	continu	e to burn	Flame not seen	-	PASS
Part 7.12 Carbondioxide content of the inhalation air	Shall not exceed an av	erage o	f % 1		0,84 0,87 0,82	-	PASS
Part 7.13 Head harness	It can be donned and	removed	deasily		Appropriate	-	PASS
Part 7.14 Field of vision	The field of vision sha performance test.	II accep	table in	practical	Appropriate	-	PASS
Part 7.15 Exhalation valve(s)	It shall withstand axia apply for 10 s. If fitted, shall contin after a continuous of L/min over a period of	ue to o	perate	correctly	Not applicable	-	Not applicable



TECHNICAL EVALUATION REPORT (162-21-01-R01)

TESTS PARAM	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE	EVALUATION
	F	FFP1	FFP2	FFP3		LEVELS	
Part 7.16 Breathing	Inhalation 30L/min	0,6 mbar	0,7 mbar	1,0 mbar	See the table below	FFP2	PASS
	Inhalation 95L/min	2,1 mbar	2,4 mbar	3,0 mbar	See the table below	FFP2	PASS
	Exhalation 160L/min	3,0 mbar	3,0 mbar	3,0 mbar	See the table below	FFP2	PASS

Breathing Resistance (mbar)	Inhalation 30L/min	Inhalation 95L/min
As recieved	0,4	1,5
As recieved	0,4	1,5
As recieved	0,3	1,5
After temperature conditioning	0,3	1,6
After temperature conditioning	0,4	1,5
After temperature conditioning	0,3	1,6
After the simulated wearing treatment	0,3	1,5
After the simulated wearing treatment	0,4	1,6
After the simulated wearing treatment	0,4	1,6

Breathing Resistance 160L/min (mbar)	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side
As recieved	2,7	2,6	2,7	2,7	2,6
As recieved	2,6	2,7	2,6	2,7	2,6
As recieved	2,7	2,6	2,6	2,7	2,6
After temperature conditioning	2,6	2,7	2,7	2,6	2,6
After temperature conditioning	2,7	2,6	2,7	2,6	2,6
After temperature conditioning	2,7	2,7	2,6	2,7	2,6
After the simulated wearing treatment	2,7	2,6	2,7	2,6	2,7
After the simulated wearing treatment	2,6	2,7	2,6	2,7	2,6
After the simulated wearing treatment	2,7	2,6	2,7	2,7	2,7



TECHNICAL EVALUATION REPORT (162-21-01-R01)

TESTS	PARAMETER	PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3			
Part 7.17 Clogging	After clogging the inhalation resistances shall not exceed. (valved)	4 mbar	5 mbar	7 mbar	Not applicable	-	Not applicable
	The exhalation resistance shall not exceed 3 mbar at 160 L/ min continuous flow. (valved)				Not applicable	-	Not applicable
	After clogging the inhalation and exhalation resistances shall not exceed. (valveless)	3 mbar	4 mbar	5 mbar	Not applicable		Not applicable
Part 7.18 Demountable part	All demountable parts (if fitted) shall be readily connected and secured were possible by hand.				Not applicable	-	Not applicable

9. DECISION PROPOSAL

Analysis and examinations ARMED AR PRO 2 model coded personal protective equipment; Respiratory Protective Devices EN 149:2001 +A1:2009- Filtered Half Masks for Protection Against Particles - Properties, Experiments and Marking standards are evaluated. It is recommended to be certified at the performance levels specified as a result of technical evaluations.

10. ATTACHMENTS

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- User Instruction

Reason for revision

: Different color products have been added.

CONTROLLER

: VOLKAN AKIN

SING

.

DATE

: 12.03.2021



EU DECLARATION OF CONFORMITY

MANUFACTURER

ASTEK METAL MEDIKAL YAPI TEKNOLOJI SANAYI TICARET LIMITED ŞİRKETİ Akburgaz Mah 3032.Sok.No:4/1 Esenyurt İstanbul

PRODUCT DESCRIPTION

Brand Name: ARMED Model :AR-PR02 Classification : FFP2 NR

Particle Filtering Half Face Mask in Category TTI product acceding to (EU) 2016/425 it is approved that the product meets the requirements of the Regulation

The Manufacturer declares on his sole responsibility that the product above is, under conditions of normal use and conditions defirmi by the Manufacturer, safe and meets all the necessary legal conditions and requirements. The product is a personel protective equipment that is intended for single use and solely in accordance with Manufacturer's instructions.

The conformity is ensured with the following mechanism:

- Complies with EU 2016/425 Person. Protective Equipment Regulation establishing technical requirements for Category III products,
- Complies with Essential Health and safety Requirements of Technical harmonized standard EN 149:2001+A1:2009
- All required test referred in above standards are conducted,
- Complies with other relevant harmonized legislation and community standarts
- For the assessment of conformity the EU Type Examination certificate (Serial No:162-21-01) is issued, after all technical evaluations for conformity to the regulation and harmonized standards. Issue date: 16.02.2021 from MNA LABORATORIES IND.TRADE.CO.LTD, as Notified Body number 2841
- The product is under surveillance of same Notified Body.nb 2841 according to the Annex III (Module B) of the PPE Regulation (EU) 2016/425, for quality assurance.

MARKING, LABELLING

Marking, labelling and user information are prepared in accordance with EU 2016/425 Personel Protective Equipment Regulation and harmonized product standards given above.

MEASURE TO ENSURE CONFORMITY

The Manufacturer declares that he has taken all necessary measures to ensure the conformity of products placed on the uo uk.with technical documentation and technical requirements for this type of product.

PLACE: Istanbul/Esenyurt

DATE: 03.03.2021

AUTHORIZED: Cihat Yıldız (General Manager)
COMPANY NAME: ASTEK METAL MEDIKAL

YAPI TEKNOLOJI SANAYI TICARET LIMITED SİRKETİ

ASTEK METAL MEDIKA API TEKNOLOJI SANATI TICARET LINITED SIRKETI Akçabulgaz Marit. 3032. Sof. No.4/1 Esenyuruts T. Tet: 0212 219 72 58 E-mail ifto@astekmedikal.com. Tic.Sic.No.226109 5 E-senyur V.D: 091 106 0876 Mersis No. 009110 352500001